



einstein™
imagine • explore • learn

Nitrate Sealed Electrode Sensor Bundle

Product Number: ENNTR-A017A



Overview

Nitrates have a wide variety of uses ranging from fertilizer to explosives. In agriculture, nitrates help form nitrogen, an essential building block for plant proteins. In explosives nitrates contain nitrogen atoms weakly bonded to oxygen. When this bond is broken, the nitrogen atoms bond strongly to other nitrogen atoms, forming very strong bonds and releasing a lot of energy. The chemical symbol for nitrate is NO_3^- .

The Nitrate Sealed Electrode Sensor can be connected to all types of einstein™ data loggers.

Typical experiments



Water Quality

- Water Quality Studies



Chemistry

- Effectiveness of Fertilizers
- Explosive materials and combustible products



- Pollution Monitoring

How it works

The Nitrate sensor contains a Permafil (non-refillable) electrode containing an oxidized form of nitrate inside a membrane. When inserted into a solution containing nitrate molecules the nitrate in the solution is attracted to the oxidized chloride in the membrane. By measuring the electrical potential of this attraction, the sensor can determine the level of nitrate in the solution. Because they only attract other nitrate molecules sealed electrode sensors work well even in solutions containing numerous elements.

Sensor specification

Concentration Range:	7×10^{-6} to 1 M (0.1 ppm to 14,000 ppm)
Resolution (12-bit):	0.15 mV
Minimum Sample Size:	5 mL in a 50 mL beaker
Default Sample Rate	10 samples per second
pH Range:	2.5 to 11 pH
Temperature Range:	0 to 50 °C
Reproducibility:	± 4%
Electrode Resistance	1 to 4 MΩ
Interfering Ions	ClO ₄ ⁻ , ClO ₃ ⁻ , I ⁻ , F ⁻

Note: Sensor cables sold separately

Contents

The Nitrate Sealed Electrode Sensor comes equipped with:

- The Nitrate Sealed Electrode Sensor
- ISE (Ion Selective Electrode) Amplifier
- 1 oz. NH₃⁻ Ionic Strength Adjuster (ISA)
- 1 oz. NH₃⁻ 10 ppm as N Standard
- 1 oz. NH₃⁻ 1000 ppm as N Standard

Solutions

ISA 2M (NH ₄) ₂ SO ₄ :	264.3 g (NH ₄) ₂ SO ₄ in 1000 mL DI water
10ppm NH ₃ ⁻ (714 mM NO ₃ ⁻):	0.06067 g NaNO ₃ in 1000 mL DI water
1000ppm NH ₃ ⁻ (0.0714 M NO ₃ ⁻):	6.067 g NaNO ₃ in 1000 mL DI water

Experimental set up

Electrode Preparation

1. Remove the protective plastic cover from the tip of the electrode and gently shake the electrode downward like a thermometer to remove any air bubbles trapped inside. Caution: Do not touch the PVC membrane with your fingers.
2. Rinse the electrode with DI water and blot dry. Do not rub dry.
3. Condition the electrode by soaking it in the provided 10 ppm NO_3^- standard solution for 30 minutes.
4. After the conditioning period, rinse the tip of the electrode with DI water and blot dry.
5. The electrode is now ready to use.

This sensor must be calibrated before use (see the **Data Logging, Calibrating and Analysis** below).

Two solutions of different concentrations (depending on the range of measurements) are used to calibrate the electrode. ISA is added to all solutions to ensure that the samples and the standards have the same ionic strength.

In addition to the aforementioned contents, you will also need:

- Wash Bottle with Distilled (DI) or deionized water.
- Several clean beakers.
- 0.1mL, 10mL pipettes.

Data logging, Calibrating and Analysis

MiLAB™ Android & IOS

1. Take your einstein™ Tablet or pair your einstein™ LabMate with your Android or iOS tablet via Bluetooth
2. Insert the electrode into the ISE amplifier
3. Insert the ISE amplifier cable into one of the sensor ports
4. Launch MiLAB
5. MiLAB will automatically detect the ISE amplifier and show it in the Launcher View
6. Tap ISE amplifier and select the Nitrate electrode



7. Make sure the icon is checked () to enable it for logging


Calibration in MiLAB™

Preparing the calibration solutions

1. Add 10 mL of the 10 ppm solution into a 50 mL beaker.
2. Add 0.2 mL of ISA and stir thoroughly.
3. Add 10 mL of the 1000 ppm solution into a 50 mL beaker.
4. Add 0.2 mL of ISA and stir thoroughly.

Calibrating the sensor

1. Tap the Settings button next to the sensor's name and tap Manual Calibration




2. Prepare the electrode as described in “Electrode preparation” above
3. Tap the “Real Reading” box of Point 1
4. Enter the value “10”
5. Rinse the electrode with DI water, blot dry and place in the beaker with the 10 ppm solution. Wait for a stable reading, and then tap the “Lock” icon 
6. Tap the “Real Reading” box of Point 2
7. Enter the value “1000”
8. Rinse the electrode with DI water, blot dry and place in the beaker with the 1000 ppm solution. Wait for a stable reading, and then tap the “Lock” icon 
9. Tap “Calibrate”
10. You are ready to run your experiment

Note: It is best to calibrate the electrode with one Real Reading below your expected reading and one Real Reading above your expected reading. For example, if you expect a reading of around 100 ppm it is best to calibrate with one Real Reading below 100 ppm and one Real Reading above 100 ppm


Note: You can prepare your own ppm solutions for calibration, using the strength of your solution as the “Real Reading”

MiLAB™ Desktop

1. Pair your einstein™ LabMate with your PC, MAC, or Linux machine via Bluetooth, or connect it via the USB cable (found in the einstein™ LabMate box).
2. Insert the electrode into the ISE amplifier
3. Insert the ISE amplifier cable into one of the sensor ports
4. Launch MiLAB
5. MiLAB will automatically detect the ISE amplifier and show it in the **Current Setup Summary** window



- Click **Full Setup**, located at the bottom of the **Current Setup Summary** window to set which ISE electrode you are using and to program the data logger's sample rate, number of samples, units of measurement, and other options




Calibrating in MiLAB™ Desktop

Calibrating in MiLAB Desktop

Preparing the calibration solutions

- Add 10 mL of the 10 ppm solution into a 50 mL beaker
- Add 0.2 mL of ISA and stir thoroughly
- Add 10 mL of the 1000 ppm solution into a 50 mL beaker
- Add 0.2 mL of ISA and stir thoroughly

1. Start MiLAB™ () and select the Nitrate electrode as described above.
2. Under the Calibrate column tap “Set” to bring up the Calibration menu



3. Prepare the electrode as described in “Electrode preparation” above.
4. Tap the “Real Reading” box of Point 1
5. Enter the value “10”
6. Rinse the electrode with DI water, blot dry and place in the beaker with the 10 ppm. Wait for a stable reading, and then click the “Lock” icon 
7. Tap the “Real Reading” box of Point 2
8. Enter the value “1000”
9. Rinse the electrode with DI water, blot dry and place in the beaker with the 1000 ppm solution. Wait for a stable reading, and then click the “Lock” icon 
10. Click “Calibrate”
11. Tap the Run button () on the main toolbar of the Launcher View to start logging

Note: It is best to calibrate the electrode with one Real Reading below your expected reading and one Real Reading above your expected reading. For example, if you expect a reading of around 100 ppm it is best to calibrate with one Real Reading below 100 ppm and one Real Reading above 100 ppm

Note: You can prepare your own ppm solutions for calibration, using the strength of your solution as the “Real Reading”

Maintenance and Electrode Storage

Short Term:

Rinse the electrode thoroughly with DI water and place the tip in a diluted standard solution (10 ppm) between measurements.

Long Term:

Rinse the electrode thoroughly with DI water, blot and store dry. Replace the cap to protect the sensing element.

Follow procedures in the sections **Electrode Preparation** before using the electrode again.

Troubleshooting

If the electrode slope is not within the normal range, the following procedure may restore the electrode.

1. Soak the electrode in the 10 ppm standard solution for 10 minutes before use.
2. Repeat the procedure outlined **Electrode Preparation** again.

Technical support

For technical support, you can contact the Fourier Education's technical support team at:

Web: www.einsteinworld.com/support

Email: support@fourieredu.com

Copyright and Warranty

All standard Fourier Systems sensors carry a one (1) year warranty, which states that for a period of twelve months after the date of delivery to you, it will be substantially free from significant defects in materials and workmanship.

This warranty does not cover breakage of the product caused by misuse or abuse.

This warranty does not cover Fourier Systems consumables such as electrodes, batteries, EKG stickers, cuvettes and storage solutions or buffers.