

EC - DECLARATION OF CONFORMITY



No.: FT65_CE_DoC, valid to 2019-07-02

Manufacturer: Beurer GmbH

Address: Söflinger Str. 218
89077 Ulm
Germany

Förderndes Mitglied der
Deutschen Hochdruckliga

Förderndes Mitglied der
Deutschen Diabetes Stiftung

This declaration of conformity is issued under our sole responsibility.

UMDNS code and name: 17-888 Thermometer, Infrared, Skin

Category code and category: 04 Electrical and mechanical medical devices

Classification Rule/ Class: Rule 10/ Class IIa

Product name: FT65

Expiry Date: 2019-07-02

The object of the declaration described above is in conformity with the relevant Community harmonisation legislation:

Directives _Date of last revision: CD 93/42/EEC (MDD)
amended with 2007/47/EEC_2010-03-21
CD 2011/65/EC (ROHS)_2011-06-08

Conformity assessment procedure: Annex II – excluding Section 4 of CD 93/42/EEC (MDD)

Relevant harmonised standards: EN 60601-1: 2006 + A11: 2011 + A1: 2013 + A12: 2014
EN 60601-1-2: 2015
EN 12470-5: 2003


The notified body mdc - medical device certification GmbH, located at Kriegerstr. 6, D-70191 Stuttgart, identification number 0483, issued in the course of conformity assessment procedure the following certificate:

Certificate No._issued on: D1311700028_2017-11-29, valid to 2019-07-02

Signed for and on behalf of: Beurer GmbH

Place, date of issue: Ulm, 2019-03-08

Name, function and signature: Werner Meternek, Director R&D/ RA


Beurer GmbH
Söflinger Str. 218 · 89077 Ulm