



einstein™
imagine • explore • learn

Ammonium Sealed Electrode Sensor Bundle

Product Number: ENAMN020A



Overview

Ammonium ions (NH_4^+) can be found in fertilizer, cleaning products, processed meats and in the waste of animals. In water, at high concentrations, it can be a dangerous pollutant. The Ammonium sensor measures the concentration of ammonium ions in aqueous solutions.

The Ammonium Sealed Electrode Sensor can be connected to the all einstein™ Tablets, einstein™Labmate™, and einstein™Labmate+™.

Typical experiments



Water Quality

- Water Quality Studies



Chemistry

- Effectiveness of Fertilizers
- Level of Ammonium in Processed Meat
- Corrosiveness of Ammonium in Canned Foods

How it works

The Ammonium sensor contains a Permafil (non-refillable) electrode containing an oxidized form of bromide inside a membrane. When inserted into a solution containing Ammonium molecules the Ammonium in the solution is attracted to the oxidized Ammonium in the membrane. By measuring the electrical potential of this attraction the sensor can determine the level of Ammonium in the solution. Because they only attract other Ammonium molecules sealed electrode sensors work well even in solutions containing numerous elements.

Sensor specification

Concentration Range:	1×10^{-6} to 1 M (0.014 - 14,000 ppm NH_4^+ as N)
Resolution (12-bit):	0.15 mV
Minimum Sample Size:	5 mL in a 50 mL beaker
Default Sample Rate	10 samples per second
pH Range:	4 to 10 pH
Temperature Range :	0 to 50 °C
Reproducibility :	$\pm 4\%$
Electrode Resistance	1 to 4 M Ω
Interfering Ions	K^+ , Na^+

Note: Sensor cables sold separately

Contents

The Ammonium Sealed Electrode Sensor comes equipped with:

- The Ammonium Sealed Electrode Sensor
- ISE (Ion Selective Electrode) Amplifier
- (1) 1 oz. NH_4^+ Ionic Strength Adjuster (ISA) (AJNH41)
- (1) 1 oz. NH_4^+ 10 ppm as N Standard (SD2052)
- (1) 1 oz. NH_4^+ 1000 ppm as N Standard (SD2002 / SDNH42)

Solutions

ISA 1M NaCl:	58.443g NaCl in 1000mL DI water
10ppm NH_4^+ as N Standard (0.0714 M NH_4^+):	38.19 mg NH_4Cl in 1000 mL DI water
1000 ppm NH_4^+ as N Standard (0.0714 M NH_4^+):	3.819 g NH_4Cl in 1000 mL DI water

Experimental set up

Electrode Preparation

1. Remove the plastic protective vial from the tip of the electrode and gently shake the electrode downward like a thermometer to remove any air bubbles trapped inside. Caution: Do not touch the PVC membrane with your fingers.

2. Rinse the electrode with DI water, blot dry. Do not rub dry.
3. Condition the electrode in the provided 10 ppm NH₄⁺ as N standard solution for 30 minutes.
4. After the conditioning period, rinse the tip of the electrode with DI water and blot dry.
5. The electrode is now ready to use.

This sensor must be calibrated before use (see the **Data Logging, Calibrating and Analysis** below).

Two solutions of different concentrations (depending on the range of measurements) are used to calibrate the electrode. ISA is added to all solutions to ensure that the samples and the standards have the same ionic strength.

In addition to the aforementioned contents you will also need:

- Wash Bottle with Distilled (DI) or deionized water.
- Several clean beakers.
- 1mL, 10mL pipettes.

Data logging, Calibrating and Analysis

MiLAB™ Android & IOS

1. Take your einstein™ Tablet or pair your einstein™LabMate™ with your Android or iOS tablet via Bluetooth
2. Insert the electrode into the ISE amplifier
3. Insert the ISE amplifier cable into one of the sensor ports
4. Launch MiLAB
5. MiLAB will automatically detect the ISE amplifier and show it in the Launcher View
6. Tap ISE amplifier and select the Ammonium electrode



7. Make sure the icon is checked () to enable it for logging


Calibration in MiLAB™



Preparing the calibration solutions

1. Add 10 mL of the 10 ppm solution into a 50 mL beaker.
2. Add 0.2 mL of ISA and stir thoroughly.
3. Add 10 mL of the 1000 ppm solution into a 50 mL beaker.
4. Add 0.2 mL of ISA and stir thoroughly.

Calibrating the sensor

1. Tap the Settings button next to the sensor's name and then Manual Calibration




2. Prepare the electrode as described in "Electrode preparation" above
3. Tap the "Real Reading" box of Point 1
4. Enter the value "10"
5. Rinse the electrode with DI water, blot dry and place in the beaker with the 10 ppm solution. Wait for a stable reading, and then tap the "Lock" icon 
6. Tap the "Real Reading" box of Point 2
7. Enter the value "1000"
8. Rinse the electrode with DI water, blot dry and place in the beaker with the 1000 ppm solution. Wait for a stable reading, and then tap the "Lock" icon 
9. Tap "Calibrate"
10. You are ready to run your experiment

Note: It is best to calibrate the electrode with one Real Reading below your expected reading and one Real Reading above your expected reading. For example if you expect a reading of around 100 ppm it is best to calibrate with one Real Reading below 100 ppm and one Real Reading above 100 ppm


Note: You can prepare your own ppm solutions for calibration, using the strength of your solution as the "Real Reading"

MiLAB™ Desktop

1. Pair your einstein™LabMate™ with your PC, MAC, or Linux machine via Bluetooth, or connect it via the USB cable (found in the einstein™LabMate™ box).
2. Insert the electrode into the ISE amplifier
3. Insert the ISE amplifier cable into one of the sensor ports
4. Launch MiLAB
5. MiLAB will automatically detect the ISE amplifier and show it in the **Current Setup Summary** window




- Click **Full Setup**, located at the bottom of the **Current Setup Summary** window to set which ISE electrode you are using and to program the data logger's sample rate, number of samples, units of measurement, and other options







Calibrating in MiLAB Desktop

Preparing the calibration solutions

- Add 10 mL of the 10 ppm solution into a 50 mL beaker
- Add 0.2 mL of ISA and stir thoroughly
- Add 10 mL of the 1000 ppm solution into a 50 mL beaker
- Add 0.2 mL of ISA and stir thoroughly

- Start MiLAB™ () and select the Ammonium electrode as described above.
- Under the Calibrate column tap "Set" to bring up the Calibration menu



3. Prepare the electrode as described in “Electrode preparation” above.
4. Tap the “Real Reading” box of Point 1
5. Enter the value “10”
6. Rinse the electrode with DI water, blot dry and place in the beaker with the 10 ppm. Wait for a stable reading, and then click the “Lock” icon 
7. Tap the “Real Reading” box of Point 2
8. Enter the value “1000”
9. Rinse the electrode with DI water, blot dry and place in the beaker with the 1000 ppm solution. Wait for a stable reading, and then click the “Lock” icon 
10. Click “Calibrate”
11. Tap the Run button () on the main toolbar of the Launcher View to start logging

Note: It is best to calibrate the electrode with one Real Reading below your expected reading and one Real Reading above your expected reading. For example if you expect a reading of around 100 ppm it is best to calibrate with one Real Reading below 100 ppm and one Real Reading above 100 ppm

Note: You can prepare your own ppm solutions for calibration, using the strength of your solution as the “Real Reading”

Maintenance and Electrode Storage

Short Term:

Rinse the electrode thoroughly with DI water and place the tip in a diluted standard solution (10 ppm) between measurements.

Long Term:

Rinse the electrode thoroughly with DI water, blot and store dry. Replace the cap to protect the sensing element.

Follow procedures in the sections **Electrode Preparation** before using the electrode again.

Troubleshooting

If the electrode reading is not within the normal range, the following procedure may restore the electrode.

1. Soak the electrode in the 10 ppm NH_4^+ as N standard solution for 2 hours before use.
2. Repeat the procedure outlined **Electrode Preparation** again.

Technical support

For technical support, you can contact the Fourier Education's technical support team at:

Web: www.einsteinworld.com/support

Email: support@fourieredu.com

Copyright and Warranty

All standard Fourier Systems sensors carry a one (1) year warranty, which states that for a period of twelve months after the date of delivery to you, it will be substantially free from significant defects in materials and workmanship.

This warranty does not cover breakage of the product caused by misuse or abuse.

This warranty does not cover Fourier Systems consumables such as electrodes, batteries, EKG stickers, cuvettes and storage solutions or buffers.

ALBERT EINSTEIN and EINSTEIN are either trademarks or registered trademarks of The Hebrew University of Jerusalem. Represented exclusively by GreenLight. Official licensed merchandise. Website: einstein.biz